



## Summary results

Sample:	HEtG 19/2-A	Assigned value:	24.769 pg/mg (Reference value)
Measurand:	EtG	Target s.d.:	6.913 pg/mg (Horwitz function)
Statistical method:	ISO 5725-5 (Alg. A+S)	Rel. target s.d.:	27.91% (Horwitz function)
Number of laboratories in calculation:	38	Range of tolerance:	10.943 - 38.595 pg/mg ( $ Z\text{-Score}  \leq 2.0$ )




## Summary results

Sample:	HEtG 19/2-B	Assigned value:	13.202 pg/mg (Reference value)
Measurand:	EtG	Target s.d.:	4.051 pg/mg (Horwitz function)
Statistical method:	ISO 5725-5 (Alg. A+S)	Rel. target s.d.:	30.68% (Horwitz function)
Number of laboratories in calculation:	36	Range of tolerance:	5.101 - 21.303 pg/mg ( $ Z\text{-Score}  \leq 2.0$ )



## Summary results

Sample:	HEtG 19/2-C	Assigned value:	59.906 pg/mg (Reference value)
Measurand:	EtG	Target s.d.:	14.639 pg/mg (Horwitz function)
Statistical method:	ISO 5725-5 (Alg. A+S)	Rel. target s.d.:	24.44% (Horwitz function)
Number of laboratories in calculation:	39	Range of tolerance:	30.629 - 89.183 pg/mg ( $ Z\text{-Score}  \leq 2.0$ )



Laboratory	HETG192A	Z score	HETG192B	Z score	HETG192C	Z score
Unit	pg/mg		pg/mg		pg/mg	
1	25.500	0.1	13.000	0.0	60.500	0.0
2	19.000	-0.8	10.000	-0.8	57.000	-0.2
3	24.000	-0.1	12.000	-0.3	56.000	-0.3
4	26.000	0.2	14.000	0.2	64.000	0.3
5	28.600	0.6	13.600	0.1	66.400	0.4
6	20.000	-0.7	9.000	-1.0	53.000	-0.5
7	18.100	-1.0	8.600	-1.1	46.500	-0.9
8	17.000	-1.1	12.000	-0.3	55.000	-0.3
9	17.400	-1.1	13.670	0.1	52.780	-0.5
10	32.000	1.0	17.000	0.9	70.000	0.7
11	13.000	-1.7	6.000	-1.8	34.000	-1.8
12	26.720	0.3	8.990	-1.0	69.600	0.7
13	< 4.000		< 4.000		47.000	-0.9
14	23.000	-0.3	10.400	-0.7	58.000	-0.1
15	32.400	1.1	16.800	0.9	75.900	1.1
16	31.300	0.9	16.800	0.9	74.200	1.0
17	19.000	-0.8	10.000	-0.8	49.000	-0.7
18	16.300	-1.2	9.860	-0.8	36.600	-1.6
19	26.100	0.2	13.100	0.0	56.100	-0.3
20	25.000	0.0	14.000	0.2	58.000	-0.1
21	12.200	-1.8			55.200	-0.3
23	27.200	0.4	14.000	0.2	62.100	0.1
24	19.000	-0.8	12.000	-0.3	44.000	-1.1
26	28.600	0.6	13.900	0.2	73.200	0.9
27	31.300	0.9	18.900	1.4	65.000	0.3
28	5.200	-2.8	2.500	-2.6	11.500	-3.3
29	10.100	-2.1	5.100	-2.0	28.200	-2.2
30	25.000	0.0	15.000	0.4	62.000	0.1
31	26.000	0.2	13.000	0.0	61.000	0.1
32	12.250	-1.8	7.200	-1.5	28.850	-2.1
33	1.435	-3.4	1.656	-2.9	33.508	-1.8
34	16.200	-1.2	7.930	-1.3	45.100	-1.0
36						
37	18.000	-1.0	11.000	-0.5	44.000	-1.1
38	16.200	-1.2	< 10.000		51.600	-0.6
39	31.000	0.9	16.000	0.7	68.000	0.6
40	23.600	-0.2	12.300	-0.2	58.000	-0.1
43	36.000	1.6	24.500	2.8	74.700	1.0
44	30.200	0.8	15.300	0.5	71.900	0.8
47	42.900	2.6	25.400	3.0	97.300	2.6